

## Transcranial direct current stimulation

### Introduction

Transcranial direct-current stimulation (tDCS) is a non-invasive form of brain stimulation similar to transcranial magnetic stimulation, but instead of using magnets, it uses a low-intensity, constant current applied through scalp electrodes. Generally, anodal stimulation induces an increase of cortical excitability, whereas cathodal stimulation decreases cortical excitability, with effects that last beyond the stimulation period. Dose involves current intensity, duration of stimulation and size of electrodes. The use of tDCS in bipolar disorder is in the early stages of investigation for relief of symptoms in people who are not satisfied with their response to medication.

### Method

We have included only systematic reviews (systematic literature search, detailed methodology with inclusion/exclusion criteria) published in full text, in English, from the year 2010 that report results separately for people with a diagnosis of bipolar or related disorders. Due to the high volume of systematic reviews we have now limited inclusion to systematic meta-analyses. Where no systematic meta-analysis exists for a topic, systematic reviews without meta-analysis are included for that topic. Reviews were identified by searching the databases MEDLINE, EMBASE, and PsycINFO. Hand searching reference lists of identified reviews was also conducted. When multiple copies of review topics were found, the most recent and/or comprehensive review was included.

Review reporting assessment was guided by the Preferred Reporting Items for Systematic Reviews and Meta-Analyses ([PRISMA](#)) checklist which describes a preferred way to present a meta-analysis<sup>1</sup>. Reviews reporting less than 50% of items have been excluded from the library. The PRISMA flow diagram is a suggested way of providing information about studies included and excluded with reasons for exclusion. Where no flow diagram has been presented by individual reviews, but identified

studies have been described in the text, reviews have been checked for this item. Note that early reviews may have been guided by less stringent reporting checklists than the PRISMA, and that some reviews may have been limited by journal guidelines.

Evidence was graded using the Grading of Recommendations Assessment, Development and Evaluation ([GRADE](#)) Working Group approach where high quality evidence such as that gained from randomised controlled trials (RCTs) may be downgraded to moderate or low if review and study quality is limited, if there is inconsistency in results, indirect comparisons, imprecise or sparse data and high probability of reporting bias. It may also be downgraded if risks associated with the intervention or other matter under review are high. Conversely, low quality evidence such as that gained from observational studies may be upgraded if effect sizes are large, there is a dose dependent response or if results are reasonably consistent, precise and direct with low associated risks (see end of table for an explanation of these terms)<sup>2</sup>. The resulting table represents an objective summary of the available evidence, although the conclusions are solely the opinion of staff of NeuRA (Neuroscience Research Australia).

### Results

We found two systematic reviews that met our inclusion criteria<sup>3, 4</sup>.

- Moderate quality evidence suggests no differences between active tDCS and sham in treatment-emergent mania or hypomania.
- Low quality evidence is unable to determine the effects of tDCS for bipolar depression.



**Transcranial direct current stimulation**

*Brunoni AR, Moffa AH, Sampaio-Junior B, Galvez V, Loo CK*

**Treatment-emergent mania/hypomania during antidepressant treatment with transcranial direct current stimulation (tDCS): A systematic review and meta-analysis**

**Brain Stimulation 2017; 10: 260-2**

[View review abstract online](#)

<b>Comparison</b>	<b>Effects of tDCS treatment for depression vs. sham on mania/hypomania outcome.</b> The studies included patients with bipolar or unipolar depression.
<b>Summary of evidence</b>	<b>Moderate quality evidence (large sample, consistent, imprecise, some indirectness) suggests no differences between active tDCS and sham in treatment-emergent mania.</b>
<b>Treatment-emergent mania/hypomania</b>	
<i>There were no significant differences between groups;</i> 10 RCTs, N = 316, OR = 1.79, 95%CI 0.6 to 5.32, $p > 0.05$ , $I^2 = 0\%$ , $p = 0.95$	
<b>Consistency in results<sup>†</sup></b>	Consistent
<b>Precision in results<sup>§</sup></b>	Imprecise
<b>Directness of results<sup>  </sup></b>	Direct comparison, indirect sample (not all patients had bipolar disorder).

*Donde C, Amad A, Nieto I, Brunoni AR, Neufeld NH, Bellivier F, Poulet E, Geoffroy PA*

**Transcranial direct-current stimulation (tDCS) for bipolar depression: A systematic review and meta-analysis**

**Progress in Neuro-Psychopharmacology & Biological Psychiatry 2017; 78: 123-31**

[View review abstract online](#)

<b>Comparison</b>	<b>Pre-post effects of tDCS + usual treatment.</b>
<b>Summary of evidence</b>	<b>Low quality evidence (very small sample, consistent, imprecise, direct) is unable to determine the effects of tDCS for bipolar depression.</b>



**Transcranial direct current stimulation**

<b>Depression</b>	
<p><i>There were no significant differences between groups;</i></p> <p>After acute phase: 7 studies, N = 46, SMD = 0.71, 95%CI 0.25 to 1.18, <math>p = 0.003</math>, <math>I^2 = 0\%</math></p> <p>At end of treatment: 7 studies, N = 46, SMD = 1.27, 95%CI 0.57 to 1.97, <math>p = 0.0004</math>, <math>I^2 = 18\%</math></p>	
<b>Risks</b>	6 cases of affective switching. There were transient, non-specific adverse effects reported including: skin redness, itching/burning sensation at the electrode site, headache, nausea, visual effects, and fatigue. No cases of rapid cycling induction or increased suicide risk were reported.
<b>Consistency in results</b>	Consistent
<b>Precision in results</b>	Imprecise
<b>Directness of results</b>	Direct

**Explanation of acronyms**

$I^2$  = the percentage of the variability in effect estimates that is due to heterogeneity rather than sampling error (chance), N = number of participants, OR = odds ratio,  $p$  = statistical probability of obtaining that result ( $p < 0.05$  generally regarded as significant), RCT = randomized controlled trial, SMD = standardized mean difference, tDCS = transcranial direct current stimulation, vs. = versus

## Transcranial direct current stimulation

### Explanation of technical terms

\* Bias has the potential to affect reviews of both RCT and observational studies. Forms of bias include; reporting bias – selective reporting of results, publication bias - trials that are not formally published tend to show less effect than published trials, further if there are statistically significant differences between groups in a trial, these trial results tend to get published before those of trials without significant differences; language bias – only including English language reports; funding bias - source of funding for the primary research with selective reporting of results within primary studies; outcome variable selection bias; database bias - including reports from some databases and not others; citation bias - preferential citation of authors. Trials can also be subject to bias when evaluators are not blind to treatment condition and selection bias of participants if trial samples are small<sup>5</sup>.

† Different effect measures are reported by different reviews.

Weighted mean difference scores refer to mean differences between treatment and comparison groups after treatment (or occasionally pre to post treatment) and in a randomised trial there is an assumption that both groups are comparable on this measure prior to treatment. Standardised mean differences are divided by the pooled standard deviation (or the standard deviation of one group when groups are homogenous) that allows results from different scales to be combined and compared. Each study's mean difference is then given a weighting depending on the size of the sample and the variability in the data. 0.2 represents a small effect, 0.5 a medium effect, and 0.8 and over represents a large treatment effect<sup>5</sup>.

Odds ratio (OR) or relative risk (RR) refers to the probability of a reduction ( $< 1$ ) or an increase ( $> 1$ ) in a particular outcome in a treatment group, or a group exposed to a risk factor, relative to the comparison group. For example, a RR of 0.75 translates to a reduction in risk of an outcome of 25% relative to those not receiving the treatment or not exposed to the risk factor. Conversely, an RR of 1.25 translates to an increased risk of 25% relative to those not receiving treatment or not having been exposed to a risk factor. An RR or OR of 1.00 means there is no difference between groups. A medium effect is considered if  $RR > 2$  or  $< 0.5$  and a large effect if  $RR > 5$  or  $< 0.2$ <sup>6</sup>. InOR stands for logarithmic OR where a InOR of 0 shows no difference between groups. Hazard ratios measure the effect of an explanatory variable on the hazard or risk of an event.

Correlation coefficients (eg,  $r$ ) indicate the strength of association or relationship between variables. They are an indication of prediction, but do not confirm causality due to possible and often unforeseen confounding variables. An  $r$  of 0.10 represents a weak association, 0.25 a medium association and 0.40 and over represents a strong association. Unstandardised ( $b$ ) regression coefficients indicate the average change in the dependent variable associated with a 1 unit change in the dependent variable, statistically controlling for the other independent variables. Standardised regression coefficients represent the change being in units of standard deviations to allow comparison across different scales.

Prevalence refers to how many existing cases there are at a particular point in time. Incidence refers to how many new cases

there are per population in a specified time period. Incidence is usually reported as the number of new cases per 100,000 people per year. Alternatively some studies present the

## Transcranial direct current stimulation

number of new cases that have accumulated over several years against a person-years denominator. This denominator is the sum of individual units of time that the persons in the population are at risk of becoming a case. It takes into account the size of the underlying population sample and its age structure over the duration of observation.

Reliability and validity refers to how accurate the instrument is. Sensitivity is the proportion of actual positives which are correctly identified (100% sensitivity = correct identification of all actual positives) and specificity is the proportion of negatives which are correctly identified (100% specificity = not identifying anyone as positive if they are truly not).

‡ Inconsistency refers to differing estimates of treatment effect across studies (i.e. heterogeneity or variability in results) which is not explained by subgroup analyses and therefore reduces confidence in the effect estimate.  $I^2$  is the percentage of the variability in effect estimates that is due to heterogeneity rather than sampling error (chance) - 0% to 40%: heterogeneity might not be important, 30% to 60%: may represent moderate heterogeneity, 50% to 90%: may represent substantial heterogeneity and 75% to 100%: considerable heterogeneity.  $I^2$  can be calculated from Q (chi-square) for the test of heterogeneity with the following formula;

$$I^2 = \left( \frac{Q - df}{Q} \right) \times 100\%$$

§ Imprecision refers to wide confidence intervals indicating a lack of confidence in the effect estimate. Based on GRADE recommendations, a result for continuous data (standardised mean differences, not

weighted mean differences) is considered imprecise if the upper or lower confidence limit crosses an effect size of 0.5 in either direction, and for binary and correlation data, an effect size of 0.25. GRADE also recommends downgrading the evidence when sample size is smaller than 300 (for binary data) and 400 (for continuous data), although for some topics, this criteria should be relaxed<sup>7</sup>.

|| Indirectness of comparison occurs when a comparison of intervention A versus B is not available but A was compared that which allows indirect comparisons of the magnitude of effect of A versus B. Indirectness of population, comparator and or outcome can also occur when the available evidence regarding a particular population, intervention, comparator, or outcome is not available so is inferred from available evidence. These inferred treatment effect sizes are of lower quality than those gained from head-to-head comparisons of A and B.



## Transcranial direct current stimulation

### References

1. Moher D, Liberati A, Tetzlaff J, Altman DG, PRISMA Group (2009): Preferred reporting items for systematic reviews and meta-analyses: the PRISMA statement. *British Medical Journal* 151: 264-9.
2. GRADE Working Group (2004): Grading quality of evidence and strength of recommendations. *British Medical Journal* 328: 1490.
3. Donde C, Amad A, Nieto I, Brunoni AR, Neufeld NH, Bellivier F, *et al.* (2017): Transcranial direct-current stimulation (tDCS) for bipolar depression: A systematic review and meta-analysis. *Progress in Neuro-Psychopharmacology & Biological Psychiatry* 78: 123-31.
4. Brunoni AR, Moffa AH, Sampaio-Junior B, Galvez V, Loo CK (2017): Treatment-emergent mania/hypomania during antidepressant treatment with transcranial direct current stimulation (tDCS): A systematic review and meta-analysis. *Brain Stimulation* 10: 260-2.
5. Cochrane Collaboration (2008): Cochrane Handbook for Systematic Reviews of Interventions. Accessed 24/06/2011.
6. Rosenthal JA (1996): Qualitative Descriptors of Strength of Association and Effect Size. *Journal of Social Service Research* 21: 37-59.
7. GRADEpro (2008): [Computer program]. Jan Brozek, Andrew Oxman, Holger Schünemann. *Version 3.2 for Windows*.